

SUPERVISOR TEST

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 11/27/2024

OPERATOR NAME:  
BENJAMIN E PETRIEE  
Permit #: 52840  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 24270

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:25
INTERNAL STANDARD	VERIFIED	17:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	17:26
BLANK TEST	0.00	17:26
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	17:27
BLANK TEST	0.00	17:28
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	17:28
BLANK TEST	0.00	17:29

WFO

