

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 12/09/2024

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:46
INTERNAL STANDARD	VERIFIED	14:47
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	14:47
BLANK TEST	0.00	14:48
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	14:48
BLANK TEST	0.00	14:49
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	14:49
BLANK TEST	0.00	14:50

