

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 01/24/2024

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:25
INTERNAL STANDARD	VERIFIED	18:25
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:25
BLANK TEST	0.00	18:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:26
BLANK TEST	0.00	18:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	18:28
BLANK TEST	0.00	18:28

