

SUPERVISOR TEST

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 03/27/2024

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52036
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:57
INTERNAL STANDARD	VERIFIED	01:57
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:57
BLANK TEST	0.00	01:58
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:59
BLANK TEST	0.00	01:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:00
BLANK TEST	0.00	02:01

