

**Infrared Instrument  
Weekly Calibration Verification Test Record**

Department: OC50

Instrument Serial #: 100197

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
1/11/24	23350	34.02	34.02	34.02	MP4192	.10	.10	.10	C. Brashear	N
1/17/24	23350	34.02	34.02	34.02	MP4192	.10	.10	.10	O. Sika	N
1/24/24	23350	34.02	34.02	34.02	MP4192	.10	.10	.10	C. Brashear	N
1/31/24	23430	34.02	RFI	RFI	MP4192	.10	RFI	RFI	O. Sika	Y
1/31/24	23430	34.02	34.02	34.02	MP4192	.10	.10	.10	O. Sika	Y
2/4/24	23430	34.02	34.02	34.02	MP4192	.10	.10	.10	Brashear	N
2/14/24	23430	34.03	34.03	34.03	MP4192	.10	.10	.10	O. Sika	N
2/21/24	23430	34.03	34.03	34.03	MP4192	.10	.10	.10	E. Hess	N
2/28/24	23430	34.03	34.03	34.03	MP4192	.10	.10	.10	O. Sika	N
5/1/24	23430	34.02	34.02	34.02	MP4192	.10	.10	.10	Brashear	Y
3/7/24	23430	34.02	34.02	34.02	MP4192	.10	.10	.10	Cusack	N
3/13/24	23430	34.03	34.03	34.03	MP4192	.10	.10	.10	O. Sika	N

\* - Please enter a ✓ to denote simulator solution change taking place

