

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 09/02/2025

OPERATOR NAME:
BRADLY P YATES
Permit #: 53896
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:25
INTERNAL STANDARD	VERIFIED	01:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	01:25
BLANK TEST	0.00	01:26
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	01:26
BLANK TEST	0.00	01:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	01:28
BLANK TEST	0.00	01:28

WFO

