

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 08/01/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:59
INTERNAL STANDARD	VERIFIED	19:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	20:00
BLANK TEST	0.00	20:01
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	20:01
BLANK TEST	0.00	20:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	20:02
BLANK TEST	0.00	20:03

CFO

