

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

WFO

Date: 04/01/2025

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:16
INTERNAL STANDARD	VERIFIED	14:16
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	14:17
BLANK TEST	0.00	14:17
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	14:18
BLANK TEST	0.00	14:19
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	14:19
BLANK TEST	0.00	14:20

