

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 12/02/2025

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:07
INTERNAL STANDARD	VERIFIED	20:07
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:07
BLANK TEST	0.00	20:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:08
BLANK TEST	0.00	20:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:10
BLANK TEST	0.00	20:10

