

WFO

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 09/19/2025

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 32057
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 52127

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:02
INTERNAL STANDARD	VERIFIED	06:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:02
BLANK TEST	0.00	06:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:04
BLANK TEST	0.00	06:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	06:05
BLANK TEST	0.00	06:06

