

SOLUTION CHANGE

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

WFO

Date: 03/01/2025

OPERATOR NAME:  
ROBERT J PRODY  
Permit #: 50912  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	23:59
INTERNAL STANDARD	VERIFIED	00:00
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:00
BLANK TEST	0.00	00:01
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	00:01
BLANK TEST	0.00	00:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	00:02
BLANK TEST	0.00	00:03

