

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

WFO

Date: 01/01/2025

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:21
INTERNAL STANDARD	VERIFIED	18:21
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:21
BLANK TEST	0.00	18:22
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:22
BLANK TEST	0.00	18:23
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	18:24
BLANK TEST	0.00	18:24

