

SOLUTION CHANGE

WFO

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 07/01/2025

OPERATOR NAME:
JASON L OSIKA
Permit #: 35567
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:11
INTERNAL STANDARD	VERIFIED	22:12
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:12
BLANK TEST	0.00	22:13
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:13
BLANK TEST	0.00	22:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:14
BLANK TEST	0.00	22:15

