

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 08/27/2025

OPERATOR NAME:
CHRISTIANN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:48
INTERNAL STANDARD	VERIFIED	05:48
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:48
BLANK TEST	0.00	05:49
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	05:49
BLANK TEST	0.00	05:50
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:50
BLANK TEST	0.00	05:51

