

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 10/01/2025

OPERATOR NAME:
BRADY A WOOD
Permit #: 52004
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:32
INTERNAL STANDARD	VERIFIED	03:32
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:33
BLANK TEST	0.00	03:34
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:34
BLANK TEST	0.00	03:35
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	03:35
BLANK TEST	0.00	03:36

