

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 12/13/2025

OPERATOR NAME:
RAYMOND Y BOYKO
Permit #: 53011
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:02
INTERNAL STANDARD	VERIFIED	08:02
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	08:02
BLANK TEST	0.00	08:03
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	08:03
BLANK TEST	0.00	08:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	08:04
BLANK TEST	0.00	08:05

WFO
Weekly

