

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 10/22/2025

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	04:46
INTERNAL STANDARD	VERIFIED	04:46
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	04:46
BLANK TEST	0.00	04:47
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	04:47
BLANK TEST	0.00	04:48
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	04:49
BLANK TEST	0.00	04:49

