

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 11/12/2025

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 552127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:39
INTERNAL STANDARD	VERIFIED	18:39
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:40
BLANK TEST	0.00	18:40
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:41
BLANK TEST	0.00	18:42
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:42
BLANK TEST	0.00	18:43

WFO

