

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 09/24/2025

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:40
INTERNAL STANDARD	VERIFIED	05:40
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:41
BLANK TEST	0.00	05:42
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:42
BLANK TEST	0.00	05:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:43
BLANK TEST	0.00	05:44

