

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 09/03/2025

OPERATOR NAME:
BENJAMIN E PETRIE
Permit #: 52840
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:32
INTERNAL STANDARD	VERIFIED	17:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:32
BLANK TEST	0.00	17:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	17:33
BLANK TEST	0.00	17:34
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:35
BLANK TEST	0.00	17:36

