

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 11/25/2025

OPERATOR NAME:
RAYMOND Y BOYKO
Permit #: 53011
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:50
INTERNAL STANDARD	VERIFIED	17:50
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:51
BLANK TEST	0.00	17:51
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:52
BLANK TEST	0.00	17:53
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:53
BLANK TEST	0.00	17:54

