

WFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 01/29/2025

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:15
INTERNAL STANDARD	VERIFIED	02:15
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:16
BLANK TEST	0.00	02:17
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:17
BLANK TEST	0.00	02:18
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:18
BLANK TEST	0.00	02:19

