

WFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 02/05/2025

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:35
INTERNAL STANDARD	VERIFIED	17:35
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:36
BLANK TEST	0.00	17:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	17:37
BLANK TEST	0.00	17:38
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	17:38
BLANK TEST	0.00	17:39

