

WFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 03/05/2025

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:48
INTERNAL STANDARD	VERIFIED	17:48
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	17:49
BLANK TEST	0.00	17:50
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.09	17:50
BLANK TEST	0.00	17:51
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	17:51
BLANK TEST	0.00	17:52

