

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 01/08/2025

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:22
INTERNAL STANDARD	VERIFIED	19:22
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:22
BLANK TEST	0.00	19:23
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:23
BLANK TEST	0.00	19:24
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	19:25
BLANK TEST	0.00	19:25

WFO

