

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 10/08/2025

OPERATOR NAME:
RAYMOND Y BOYKO
Permit #: 53011
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:22
INTERNAL STANDARD	VERIFIED	22:22
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:23
BLANK TEST	0.00	22:23
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:24
BLANK TEST	0.00	22:25
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	22:25
BLANK TEST	0.00	22:26

WFO Weekly

