

WFO

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 12/17/2025

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25330

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:52
INTERNAL STANDARD	VERIFIED	02:52
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:52
BLANK TEST	0.00	02:53
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:54
BLANK TEST	0.00	02:54
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	02:55
BLANK TEST	0.00	02:56

