

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 10/15/2025

OPERATOR NAME:
BENJAMIN E PETRIE
Permit #: 52840
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25250

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:45
INTERNAL STANDARD	VERIFIED	17:45
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:45
BLANK TEST	0.00	17:46
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	17:47
BLANK TEST	0.00	17:47
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	17:48
BLANK TEST	0.00	17:49

