

SUPERVISOR TEST

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100197

WFO

Date: 01/16/2025

OPERATOR NAME:
 STEVEN P EILERS
 Permit #: 49321
 Operator Agency/Dept: ONEIDA CO. S.O.
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:51
INTERNAL STANDARD	VERIFIED	14:52
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	14:52
BLANK TEST	0.00	14:53
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.09	14:53
BLANK TEST	0.00	14:54
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	14:54
BLANK TEST	0.00	14:55

