

SUPERVISOR TEST

ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100197

Date: 01/22/2025

OPERATOR NAME:
 STEVEN EILERS
 Permit #: 49321
 Operator Agency/Dept: ONEIDA CO. S.O.
 Reference Standard #: 24350

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:53
INTERNAL STANDARD	VERIFIED	18:53
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:54
BLANK TEST	0.00	18:54
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:55
BLANK TEST	0.00	18:56
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:56
BLANK TEST	0.00	18:57

