

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 02/26/2025

WFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24410

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 18:19 |
| INTERNAL STANDARD | VERIFIED | 18:19 |
| REFERENCE STANDARD TEMP | 34.03c | |
| REFERENCE STANDARD | 0.09 | 18:20 |
| BLANK TEST | 0.00 | 18:21 |
| REFERENCE STANDARD TEMP | 34.04c | |
| REFERENCE STANDARD | 0.09 | 18:21 |
| BLANK TEST | 0.00 | 18:22 |
| REFERENCE STANDARD TEMP | 34.04c | |
| REFERENCE STANDARD | 0.10 | 18:22 |
| BLANK TEST | 0.00 | 18:23 |

