

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 03/26/2025

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:56
INTERNAL STANDARD	VERIFIED	17:56
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	17:57
BLANK TEST	0.00	17:57
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	17:58
BLANK TEST	0.00	17:59
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.10	17:59
BLANK TEST	0.00	18:00

