

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 04/16/2025

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:17
INTERNAL STANDARD	VERIFIED	18:17
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:18
BLANK TEST	0.00	18:18
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:19
BLANK TEST	0.00	18:20
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:20
BLANK TEST	0.00	18:21

