

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

WFO

Date: 04/02/2025

OPERATOR NAME:
ROBERT J PRODY
Permit #: 50912
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:27
INTERNAL STANDARD	VERIFIED	20:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	20:28
BLANK TEST	0.00	20:29
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	20:29
BLANK TEST	0.00	20:30
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:30
BLANK TEST	0.00	20:31

