

SUPERVISOR TEST

 ONEIDA COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 100197

Date: 04/25/2025

OPERATOR NAME:
 JASON L OSIKA
 Permit #: 35567
 Operator Agency/Dept: ONEIDA CO. S.O.
 Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:32
INTERNAL STANDARD	VERIFIED	21:32
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	21:32
BLANK TEST	0.00	21:33
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	21:34
BLANK TEST	0.00	21:34
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	21:35
BLANK TEST	0.00	21:36

