

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 04/30/2025

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	22:13
INTERNAL STANDARD	VERIFIED	22:14
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:14
BLANK TEST	0.00	22:15
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:15
BLANK TEST	0.00	22:16
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	22:16
BLANK TEST	0.00	22:17

