

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 05/28/2025

WPA

OPERATOR NAME:
STEVEN P EILERS
Permit #: 40321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:42
INTERNAL STANDARD	VERIFIED	01:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	01:43
BLANK TEST	0.00	01:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	01:44
BLANK TEST	0.00	01:45
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.09	01:45
BLANK TEST	0.00	01:46

