

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197
Date: 05/07/2025

OPERATOR NAME: STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25040

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:05
INTERNAL STANDARD	VERIFIED	20:05
REFERENCE STANDARD TEMP	34.03c	20:06
REFERENCE STANDARD	0.09	20:06
BLANK TEST	0.00	20:06
REFERENCE STANDARD TEMP	34.03c	20:07
REFERENCE STANDARD	0.10	20:07
BLANK TEST	0.00	20:08
REFERENCE STANDARD TEMP	34.03c	20:08
REFERENCE STANDARD	0.09	20:08
BLANK TEST	0.00	20:09

