

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 06/11/2025

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:53
INTERNAL STANDARD	VERIFIED	17:53
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:54
BLANK TEST	0.00	17:54
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:55
BLANK TEST	0.00	17:56
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:56
BLANK TEST	0.00	17:57

WFO

