

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 06/18/2025

OPERATOR NAME:
CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	06:09
INTERNAL STANDARD	VERIFIED	06:09
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	06:10
BLANK TEST	0.00	06:11
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.09	06:11
BLANK TEST	0.00	06:12
REFERENCE STANDARD TEMP	34.04c	
REFERENCE STANDARD	0.09	06:12
BLANK TEST	0.00	06:13

