

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 06/06/2025

WFO

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:42
INTERNAL STANDARD	VERIFIED	17:42
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:42
BLANK TEST	0.00	17:43
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:44
BLANK TEST	0.00	17:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	17:45
BLANK TEST	0.00	17:46

