

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 07/23/2025

OPERATOR NAME:
BENJAMIN E PETRIE
Permit #: 52840
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:07
INTERNAL STANDARD	VERIFIED	18:07
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:08
BLANK TEST	0.00	18:08
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:09
BLANK TEST	0.00	18:10
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	18:10
BLANK TEST	0.00	18:11

