

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATA MASTER dmt: 100197
Date: 07/09/2025

OPERATOR NAME: CHRISTIAN T BRASHEAR
Permit #: 52127
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25110

Nominal = 0.10 g/210L

BLANK TEST	0.00	18:29
INTERNAL STANDARD	VERIFIED	18:29
REFERENCE STANDARD TEMP	34.03c	18:30
REFERENCE STANDARD	0.09	18:31
BLANK TEST	0.00	18:31
REFERENCE STANDARD TEMP	34.03c	18:32
REFERENCE STANDARD	0.09	18:32
BLANK TEST	0.00	18:33

