

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

WFO

Date: 08/06/2025

OPERATOR NAME:
ROBERT J PRODY
Permit #: 51902
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 25170

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:27
INTERNAL STANDARD	VERIFIED	19:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:28
BLANK TEST	0.00	19:29
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	19:29
BLANK TEST	0.00	19:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	19:30
BLANK TEST	0.00	19:31

