

SUPERVISOR TEST

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

WFO

Date: 02/12/2025

OPERATOR NAME:
ROBERT J RODY
Permit #: 50912
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 24410

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:27
INTERNAL STANDARD	VERIFIED	00:27
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:28
BLANK TEST	0.00	00:28
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:29
BLANK TEST	0.00	00:30
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:30
BLANK TEST	0.00	00:31

