

WFO

Infrared Instrument  
Weekly Calibration Verification Test Record

Department: OCFO

Instrument Serial #: 100197

Test Date	Simulator Solution Lot #	Simulator Temperature			Sim. Serial#	Test Results			Breath Analysis Supervisor	Sim. Sol. Change *
		1	2	3		1	2	3		
10/31/2024	24270	34.03	34.03	34.03	MP4192 <del>100197</del>	.09	.09	.09	M. Cadwell	No
11/04/2024	24270	34.03	34.03	34.03	MP4192 <del>100197</del>	.10	.10	.10	C. Brashear	Yes
11/14/2024	24270	34.03	34.03	34.03	MP4192 <del>100197</del>	.10	.10	.10	C. Brashear	No
11/20/24	24270	34.03	34.03	34.03	MP4192 <del>100197</del>	.09	.10	.10	O Sika	NO
11/27/24	24270	34.03	34.04	34.04	MP4192	.10	.10	.10	Petrie	NO
12/2/24	24350	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	Yes
12/9/24	24350	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	No
12/11/24	24350	34.03	34.03	34.03	MP4192	.09	.09	.10	Eilers	NO
12/17/24	24350	34.04	34.03	34.04	MP4192	.10	.10	.10	O Sika	NO
12/25/24	24350	34.03	34.03	34.04	MP4192	.10	.09	.09	O Sika	NO
1/1/24	24350	34.03	34.03	34.03	MP4192	.10	.10	.10	Eilers	Yes
1/8/24	24350	34.03	34.03	34.03	MP4192	.10	.10	.10	Brashear	NO

\* - Please enter a √ to denote simulator solution change taking place

WV

100

100

100

100

100

100

100

100