

# SOLUTION CHANGE

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 08/02/2026

OPERATOR NAME:  
BENJAMIN E PETRIE  
Permit #: 52840  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 26190

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 06:20 |
| INTERNAL STANDARD       | VERIFIED | 06:20 |
| REFERENCE STANDARD TEMP | 34.03c   |       |
| REFERENCE STANDARD      | 0.09     | 06:21 |
| BLANK TEST              | 0.00     | 06:21 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 06:22 |
| BLANK TEST              | 0.00     | 06:23 |
| REFERENCE STANDARD TEMP | 34.02c   |       |
| REFERENCE STANDARD      | 0.10     | 06:23 |
| BLANK TEST              | 0.00     | 06:24 |

