

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 09/09/2026

OPERATOR NAME:
STEVEN P EILERS
Permit #: 49321
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26190

Nominal = 0.10 g/210L

BLANK TEST	0.00	21:30
INTERNAL STANDARD	VERIFIED	21:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:31
BLANK TEST	0.00	21:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	21:32
BLANK TEST	0.00	21:33
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	21:33
BLANK TEST	0.00	21:34

WFO

