

SOLUTION CHANGE

ONEIDA COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 100197

Date: 04/01/2026

OPERATOR NAME:
BAILEY J DIXON
Permit #: 50814
Operator Agency/Dept: ONEIDA CO. S.O.
Reference Standard #: 26050

WFO

Nominal = 0.10 g/210L

BLANK TEST	0.00	20:04
INTERNAL STANDARD	VERIFIED	20:04
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	20:04
BLANK TEST	0.00	20:05
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:05
BLANK TEST	0.00	20:06
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	20:06
BLANK TEST	0.00	20:07

