

WFO

SOLUTION CHANGE

ONEIDA COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 100197

Date: 02/01/2026

OPERATOR NAME:  
CHRISTIAN T BRASHEAR  
Permit #: 52127  
Operator Agency/Dept: ONEIDA CO. S.O.  
Reference Standard #: 25390

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:46
INTERNAL STANDARD	VERIFIED	05:46
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:46
BLANK TEST	0.00	05:47
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:48
BLANK TEST	0.00	05:48
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.10	05:49
BLANK TEST	0.00	05:50

